

Derived Quantities

LSI LASTEM data loggers have an internal library of formulas for calculating a series of environmental quantities. These quantities are obtained from measured parameters, entered by the operator and other calculated quantities.

Derived Quantities	Manual inputs by operator	Automatic inputs from measurements	E-LOG ALIEM	M-LOG (ELO009)
Absolute Humidity (VDI3786 Part 4)		- Temperature (°C) - Relative Humidity (%)	X	X
Air changes "1997 Ashrae Fundamentals Handbook"- fluid flow	- Room volume (m³) - Pipe dimensions (cm²) - Pipe factor	Air speed (m/s)	X	X
Air flow (mass, volume) "1997 Ashrae Fundamentals Handbook"- fluid flow	- Pipe dimension (cm²) - Pipe factor	- Air speed (m/s) - Temperature (°C) - Pressure (hPa) for mass air flow only (kg/hr)	X	X
Air speed (using Pitot and darcy tubes)	Pitot or Darcy constant value	- Differential pressure (hPa) - Absolute pressure (hPa) - Air temperature (°C)	X	X
Absolute Pressure (sea level)	Altitude	- Absolute pressure (hPa) - Temperature (°C) - Relative Humidity (%)	X	X
Dew Point (ISO7726)		- Temperature (°C) - Relative Humidity (%)	X	X
Heat index		- Temperature (°C) - Relative Humidity (%)	X	X
Evaporation	Maximum water level (mm)	Water level (mm)	X	
Mean Radiant Temperature (ISO7726)		- Radiant Temperature (°C) - Air temperature (°C) - Air speed (m/s)	X	X
Radiant temperature asymmetry (ISO7726)		- Radiant Temperature (°C) - Air temperature (°C) - Air speed (m/s)	X	X
Operative Temperature (ISO7726)		- Radiant Temperature (°C) - Air temperature (°C)	X	X
Plane Radiant Temperature 1&2 (ISO7726)	Sensor orientation (cold/warm wall, cold/warm floor)	- Net radiation (W/m²) - Sensor temperature (°C)	X	X
Partial vapour pressure (ISO7726 : Table D.1 Humidity conver.equations)		- Temperature (°C) - Relative Humidity (%)	X	X
Psychrometric Humidity (ISO7726)	- Psychrometric coefficient - Standard pressure (if measurement of absolute pressure is not present)	- Dry bulb temperature (°C) - Wet bulb temperature (°C) - Absolute pressure (hPa) (if present)	X	X

Derived Quantities	Manual inputs by operator	Automatic inputs from measure- ments	E-LOG ALIEM	M-LOG (ELO009)
Specific Humidity (VDI3786 Part 4)		- Temperature (°C) - Relative Humidity (%) - Absolute pressure (hPa)	X	X
Sunshine Duration	Sunshine threshold	Direct radiation (W/m ²)	X	X
TCH - Chill temperature (ISO TR11079)		- Temperature (°C) - Air speed (m/s)	X	X
Wet bulb Temperature (ISO7726)		- Temperature (°C) - Relative Humidity (%) - Absolute pressure (hPa)	X	X
UV exposition		- UV-A radiation (W/m²) - UV-B radiation (W/m²)	X	X
UV index		- UV-A radiation (W/m²) - UV-B radiation (W/m²)	X	X
Wind Chill index (ISO TR 11079)		- Temperature (°C) - Air speed (m/s)	X	X
WBGT index IN+OUT formulas (ISO7243)		- Outdoor air temperature (°C) - Black globe temperature (°C) - Wet temperature natural ventilation (°C)	X	X
% dissatisfied by floor temperature (ISO7730)		Floor temperature (°C)	X	X
% dissatisfied by vertical temperature (ISO7730)		- Ankle air temperature @H.10 cm (°C) - Air temperature @H.110 cm (°C)	X	X
% dissatisfied by radiant asymetry (ISO7730)	Sensor orientation (cold/warm wall, cold/warm floor)	- Net radiation (W/m²) - Sensor temperature (°C)	X	X
% dissatisfied by air draw (ISO7730)		- Turbulence index (TU using ESV306) - Air temperature (°C) - Air speed (m/s)	X	X
Daylight factor (IESNA Lighting handbook)		- Indoor lighting (lx) - Outdoor lightning (lx)	X	X
Light intensity	Distance from luminous source	Lightning (lx)	X	X
Estimated Natural Decay of SARS-CoV-2 on surfaces (stainless steel and ABS plas- tic) under a range of tempe- ratures and relative humidity (proposed by DHS and S&T, USA)	Decay factor K: 1 (50%) 13,29 (99.99%) 19,94 (99.9999%) 26,58 (99.999999%)	- Temperature (°C) - Relative Humidity (%)		X

Derived Quantities	Manual inputs by operator	Automatic inputs from measurements	E-LOG ALIEM	M-LOG (ELO009)
Math formulas				
Sum/Difference/ Multiplication/Division			X	X
Difference between two previous values			X	X
Integral/Mean/Elevation to power/Exponential/ Logarithm/ Base 10 logarithm/Square root			X	X
Mobile Maximum/Mobile Minimum/Mobile Average/ Mobile Total			X	X
Vectorial calculations specific for anemometric quantities (WMO N°8, cap. 5.8)				
Average direction (Prevailing)		Wind direction	X	X
Resulting average direction		- Wind direction - Wind speed	X	X
Resulting average speed		- Wind direction - Wind speed	X	X
Direction standard deviation		Wind direction	X	X
Calm wind percentage		Wind speed	X	X

Alpha-Log

Derived Quantities	Manual inputs by operator	Automatic inputs from measurements
Dew point (ISO 7726)		- Air temperature (°C) - Relative humidity (%)
Psychrometric Humidity (ISO7726)	- Psychrometric coefficient - Standard pressure (if measurement of absolute pressure is not present)	- Dry bulb temperature (°C) - Wet bulb temperature (°C) - Absolute pressure (hPa)
Sun position (height and azimuth) ("Astronomical Algorithms, Willmann-Bell, Inc., Richmond, VA" Book by Jean Meeus)	- Latitude - Longitude - Date/time	
Diffuse solar radiation (WMO - Guide Nr. 8, Chapt. 8, eq. 8.2)		- Global solar radiation (W/m²) - Direct solar radiation (W/m²) - Sun elevation angle (deg) (calculated)
Sunshine presence (WMO - Guide Nr. 8, Chapt. 8.1)		Direct solar radiation (W/m ²)
Sunshine presence (Step Algorithm, "Updating and development of methods for worldwide accurate measurements of sunshine duration - Vuerich, Morel, Mevel, Oliviéri")		- Global solar radiation (W/m²) - Sun elevation angle (deg) (calculated)
Sunshine presence (MeteoFrance Algorithm, "Updating and development of methods for worldwide accurate measurements of sunshine duration - Vuerich, Morel, Mevel, Oliviéri")	- Coeff. A - Coeff. B	- Global solar radiation (W/m²) - Sun elevation angle (deg) (calculated) - Date/time
Daily Evapo-transpiration (Penman- Monteith eq., "Crop evapotranspiration - Guidelines for computing crop water requirements - FAO Irrigation and drainage paper 56", Chapt. 2)	- Latitude - Altitude (m) - Height of the wind velocity measurement (m) - Global solar radiation (W/m²) (instead of Net solar radiation if absent)	- Air temperature (°C) - Relative humidity (%) - Wind velocity (m/s) - Net solar radiation (W/m²) (if present)
Barometric pressure (QNH, ref. ICAO) (Tab. 3.10 from "International Meteorological Tables - WMO No. 188 TP. 94 - 1966")	- Coeff. A - Coeff. B	Atmospheric Pressure (hPa)
Barometric pressure (QNH ref. ICAO) (ISA eq. 7, CIMO, "CIMO/ET-Stand-1/Doc. 10 (20.XI.2012)")	Altitude (m)	Atmospheric Pressure (hPa)
Barometric pressure (QNH ref. ICAO) (ICAO eq. 28/29, "CIMO/ET-Stand-1/Doc. 10 (20.XI.2012); ICAO documents 7488/9837")	Altitude (m)	Atmospheric Pressure (hPa)
WBGT index IN+OUT formulas (ISO7243)		- Outdoor air temperature (°C) - Black globe temperature (°C) - Wet temperature natural ventilation (°C)

Alpha-Log

Derived Quantities	Manual inputs by operator	Automatic inputs from measurements
Cross Wind (ICAO doc. 8896, AN/893)	Runway angle respect to the North (°)	- Wind velocity (m/s) - Wind direction (m/s)
Head Wind (ICAO doc. 8896, AN/893)		
Tail wind (ICAO doc. 8896, AN/893)		
Stability category Pasquill-Gifford (SRDT method, "Bowen, Dewart, Chen, 1983. Stability Class Determination: A Comparison for one site. Proceedings 6 th Symposium on Turbulence and Diffusion, American Meteorological Society")	Radiation threshold day/night (W/m ²)	- Wind velocity (m/s) - Solar radiation (W/m²) - Lower air temperature (°C) - Upper air temperature (°C)
Gravimetric water content (VDI 3477 - Biological waste gas purification - Biofilters)	Density of the dry part (kg/dm ³)	Fraction of gravimetric water content (-)
Universal Thermal Climate Index (UTCI) (International Society of Biometeorology ISB)		- Air temperature (°C) - Dew point (or relative humidity) (°C) - Wind speed (m/s) - Mean radiant temperature (°C)
Math formulas		
Sum/Difference/Multiplication/Division		
Difference between two previous values		
Integral/Mean/Elevation to power/ Exponential/Logarithm/ Base 10 logarithm/Square root		
Mobile Maximum/Mobile Minimum/Mobile Average/Mobile Total		
Vectorial calculations specific for anemometric quantities (WMO N°8, cap. 5.8)		
Average direction (Prevailing)		Wind direction
Resulting average direction		- Wind direction - Wind speed
Resulting average speed		- Wind direction - Wind speed
Direction standard deviation		Wind direction
Calm wind percentage		Wind speed